


Description List:

Pos.	Qty.	Part description	Reference	Approved Material
1	1	Silicone membrane	Memscap VS32069	Wacker silicone SILPURAN 6000/60
2	1	844 Dome house	Memscap M30168 2 cavity mold, Memscap Inr. 7984 4 cavity mold, Memscap Inr. 7983	Polycarbonate Bayer Makrolon, 2458
3	1	Luer rotating nut	Elcam 531008	Makrolon RX1805 Transparent PC
4	1	Protection Cap Memscap	Memscap M30174	Terluran GP-22, ABS Colour code: RAL 6018



Part Description, details

Material:		Date	Sign.	General Tolerance:	00	074A/02	-	-
According to Description List					Rev	Ark ref/EM	Date	Doc
Memscap AS	Author	150702	PTN	844-28 Dome sterile	Parts and Assembly		Rev	
	Checked	-			M30144		01	
	Approved	-						



Assembled Product, 844-28 Dome sterile. Part no.: 30144

Rev.	EM	Date	Responsible	Description
01	1933/17	2017-06-15	RS	Drawing removed. Updated with detailed description list and material information