



Assembled product P/N30678

Pos.	Qty.	Part description	Reference	Approved material
1	1	Protection Cap	Memscap, M30174	Terturan GF-22 ABS
2	1	Silicone Membrane	Memscap, VS32069	Wacker silicone, SILPURAN 6000/60
3	1	Luer rotating nut	Eicam, Part no.: 531008	Makrolon RK1805, Transparent
4	1	Dome house	Memscap, VS30168	Polycarbonate, Bayer Makrolon 2458
5	1	Female Luer lock, 3 mm	Lucomed, 9700200000	PVC Phthalate free
6	1	PVC-tube 1,5*3mm	Moden plast	PVC RB1 / DUNING
7	1	Male Luer lock	Lucomed or Hoerotronnic	ABS Terluc 2802 HD

01

01

Changed Date	Designed By	Checked Date	Checked By	Approved Date	Approved By	Changed Description/Reference	Revision
06.04.2017	RS			06.04.2017	SST	EM 1913-17 Added material list and protection cap (ref. pos 1)	01
05.05.2006				27.04.2006		ArtNr 365-05 1. st. edition	00
				05.05.2006			

  

Customer Description	30678	Description	844-33, Dome Sterile	Document No.	M30678	Revision	01
Customer Description		Surface Treatment		Document type		Project No.	